

View Lab Monthly Summary

Lab Name Govt. PHC, Å Rayappanpatti
 Lab No 12718
 Month June
 Year 2023
 Constituent Group Chemistry III

[Details About Robust Analysis](#)
[Detail About Monthly Summary](#)
[Detail about SDI](#)

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 21/06/2023

Date of Report Published : 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	266 mg/dL	-1.39	0.69
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	192	32.69	12.96	4.24	34 mg/dL	0.31	0.61
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	1.8 mg/dL	-4.53	0.01
4	URIC ACID III	Enzymatic / Uricase Colorimetric III	Robonik (Semi automatic / Fully Automatic)	34	6.25	4.51	0.28	4.2 mg/dL	-7.32	0.10
5	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	114 mg/dL	0.05	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within Å±1.01 to Å±2.00	Good.
Within Å±2.01 to Å±2.99	Accept with caution. Warning Signal.
Beyond Å±3.0	Unacceptable performance. Action Signal.

Links

[Modify Address](#)
[Yearly Summary](#)

[Modify Methods](#)
[Change Password](#)

[Submit Lab Results](#)
[Home](#)

[View Lab Monthly Summary](#)
[Sign-out](#)