

**CALIBRATION CERTIFICATE: PRIETEST TOUCH PLUS**

Customer Name	The Medical Officer	Model	Prietest Touch Plus
Address	CTM Urban Primary Health Centre	S No	ATP0600122RBK
	Srinivasapuram	Date of Calib	31-01-2023
	Coimbatore	Due date Calib	30.01.2024
		Calibrated By	Krishna Kumar K

UPS Power Supply	Observed Reading	Remarks
Earth To Neutral ( 0 - 3V AC )	0VAC	
Earth to Phase ( 225 - 230VAC )	226VAC	Measured Voltages are within the specifications
Phase to Neutral ( 225 - 230VAC )	225VAC	

Measuring Unit :	Expected Reading	Observed Reading	Remarks
Lamp Voltage	6 v	6.01v	No adjustments

	Good/ No Good	Remarks
Pump	Good	
Filterwheel motor	Good	Performance of the parts are in well within working specifications.
Flowcell condition	OK	
Lens check	OK	

**Filters Check:**

Wavelength	Film type	Range	Absorbance	Obtained
340	HP-298	0.300-0.330		0.315
	2DP- 210	2.215 -2.460		2.292
	TP -211	1.587-1.721		1.624
405	2DP- 210	1.262-1.354		1.276
	TP-211	0.976-1.044		0.987
	2TP-211	1.943-2.101		1.999
505	2DP- 210	1.220-1.310		1.262
	TP-211	0.942-1.044		0.974
	2TP-211	1.871-2.037		1.957
545	SP- 209	0.298-0.320		0.309
	2DP-210	1.126-1.198		1.161
	2TP-211	1.746-1.878		1.824
578	2DP- 210	1.095-1.164		1.123
	TP-211	0.845-0.900		0.867
	3TP-211	2.482-2.673		2.618
630	2DP- 210	1.280-1.382		1.332
	TP-211	1.015-1.119		1.051
	2TP-211	2.021-2.184		2.122

No-3, 1st Flr, Krishnakamalam Pride  
rathiar Rd, Coimbatore-641044, India

+91 422 4272725

info@greenmedicare.com

www.greenmedicare.com

GST No. 33AALFG1685J1Z6