



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JULY 2023



PC-1024

Lab Name **Govt. Block PHC, Ottanatham** Lab No **7561**  
 Constituent Group **Chemistry III** Date of Result Entered : **18/07/2023**  
 PT item **Lyophilized human serum based** Date of Report Published : **07/08/2023**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	108.7 mg/dL	-0.11	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1850	5.83	6.00	0.35	5.68 mg/dL	-0.43	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.66	114.9 mg/dL	0.41	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

**LAB ADDRESS :**

Govt. Block PHC, Ottanatham  
 THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, THOOTHUKUDI DISTRICT.  
 Ottanatham  
 TAMILNADU628302

Coordinator Contact Details:  
 Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
 Contact Number: 0416-2283102

*Pamela Christudoss*  
 Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

\*\*\*\*\* End of Report \*\*\*\*\*