

PHC, Kadalaiyur

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

st

istry III

Click on the analyte to view Graphical Data

All Analyser Result

Date of Result Entered : 11/08/2023

Date of Report Published : 01/09/2023

Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
				CV	SD			
POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	178 mg/dL	0.27	0.42
es Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.57 mg/dL	-1.60	0.01
D-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	115.9 mg/dL	0.45	0.22

Interpretation
Excellent.
Good.
Accept with caution. Warning Signal.
Unacceptable performance. Action Signal.

D. Ganesan
 MEDICAL OFFICER
 GOVT. PRIMARY HEALTH CENTRE
 KADALAIYUR