



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
 CMC EXTERNAL QUALITY ASSURANCE SCHEME  
 MONTHLY SUMMARY REPORT - SEPTEMBER 2023



Lab Name: Govt. UPHC, Gandhiji road      Lab No: 10705  
 Constituent Group: Chemistry III      Date of Result Entered: 19/09/2023  
 PT item: Lyophilized human serum based      Date of Report Published: 05/10/2023

Sl No	Analyte	Reagent	Instrument	Ref. Range	DV	CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1870	310.86	4.82	14.98	316 mg/dL	0.34	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.29	8.14	0.10	1.29 mg/dL	0.00	0.00
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1783	113.62	3.80	4.32	113 mg/dL	-0.14	0.20

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

CMC EQAS does not sub contract any components \*\*\*\*\* End of Report \*\*\*\*\*

LAB ADDRESS :  
 Govt. UPHC, Gandhiji road  
 MEDICAL OFFICER, GANDHIJI ROAD, ERODE CORPORATION.  
 Gandhiji road  
 TAMILNADU638102

Coordinator Contact Details:  
 Email:clinqc@cmcvellore.ac.in  
 Contact Number: 0416-2283102

*Pamela Christudoss*  
 Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential  
 CMC EQAS does not sub contract any components  
 \*\*\*\*\* End of Report \*\*\*\*\*

*Shun*  
 MEDICAL OFFICER  
 GANDHIJI ROAD URBAN  
 PRIMARY HEALTH CENTRE  
 ERODE CORPORATION