



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

Lab Name Govt. PHC, Paratharami Lab No 13033  
Constituent Group Chemistry III Date of Result Entered : 19/06/2023  
PT item Lyophilized human serum based Date of Report Published : 04/07/2023

| Sl.No | Analyte        | Method / Principle Name               | Analyzer Name                                | No of Participants | DV     | Participants |       | Your Value | SDI   | U    |
|-------|----------------|---------------------------------------|--|--------------------|--------|--------------|-------|------------|-------|------|
|       |                |                                       |  |                    |        | CV           | SD    |            |       |      |
| 1     | GLUCOSE III    | GOD-POD III                           | Robonik ( Semi automatic / Fully Automatic ) | 1823               | 286.63 | 5.17         | 14.82 | 320 mg/dL  | 2.25  | 0.69 |
| 2     | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik ( Semi automatic / Fully Automatic ) | 1848               | 2.57   | 6.64         | 0.17  | 2.3 mg/dL  | -1.58 | 0.01 |
| 3     | CHOLESTROL III | CHOD-PAP III                          | Robonik ( Semi automatic / Fully Automatic ) | 1804               | 113.73 | 4.46         | 5.07  | 122 mg/dL  | 1.63  | 0.24 |

| SDI Range                       | Interpretation                           |
|---------------------------------|--|
| Within -1.00 to +1.00           | Excellent.                               |
| Within $\pm 1.01$ to $\pm 2.00$ | Good.                                    |
| Within $\pm 2.01$ to $\pm 2.99$ | Accept with caution. Warning Signal.     |
| Beyond $\pm 3.0$                | Unacceptable performance. Action Signal. |

LAB ADDRESS :  
Govt. PHC, Paratharami  
PARADARAMI  
Paratharami  
TAMILNADU632603

Coordinator Contact Details:  
Email:clinqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.-  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*