

**VIEW LAB MONTHLY SUMMARY**

<b>Lab Name</b>	Govt. PHC, Paramankurichi		<a href="#">Details About Robust Analysis</a>							
<b>Lab No</b>	12990	<a href="#">Detail About Monthly Summary</a>								
<b>Month</b>	September	<a href="#">Detail about SDI</a>								
<b>Year</b>	2023	<a href="#">All Analyser Result</a> <a href="#">Print</a>								
<b>Constituent Group</b>	Chemistry III	<a href="#">Click on the analyte to view Graphical Data</a>								
<b>Date of Result Entered : 20/09/2023</b>										
<b>Date of Report Published : 05/10/2023</b>										
Sl.No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1870	310.86	4.82	14.98	308 mg/dL	-0.19	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.29	8.14	0.10	1.24 mg/dL	-0.48	0.00
<b>SDI Range</b>		<b>Interpretation</b>								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

**Links**

[Modify Address](#)

[Modify Methods](#)

[Submit Lab Results](#)

[View Lab Monthly Summary](#)