

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Paramankurichi		Details About Robust Analysis	
Lab No	12990		Detail About Monthly Summary	
Month	August		Detail about SDI	
Year	2023			
Constituent Group	Chemistry III		All Analyser Result Print	
Click on the analyte to view Graphical Data				
Date of Result Entered :	19/08/2023			
Date of Report Published :	01/09/2023			

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants		Participants		Your Value	SDI	U
				Participants	DV	CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	174 mg/dL	-0.18	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.78 mg/dL	0.17	0.01
SDI Range										
Interpretation										
Within -1.00 to +1.00										
Excellent.										
Within ±1.01 to ±2.00										
Good.										
Within ±2.01 to ±2.99										
Accept with caution. Warning Signal.										
Beyond ±3.0										
Unacceptable performance. Action Signal.										

19/08/23
 174 mg/dL

Links

Modify Address

Modify Methods

Submit Lab Results

View Lab Results

மருத்துவ அலுவலர்
 அரசு ஆராய்ச்சி கலாநாயகர் நிலையம்
 புறமலை குறிச்சி - 608 213
 View Lab Results