
AMS Healthcare

(Erstwhile: Ace Medical Solutions (India) Pvt. Ltd.)

An ISO 9001-2015 Certified Company





GSTIN. 33AOSPA9153Q1ZA • DL No. TN-12-20B -00398, TN-12-21B-00398

CALIBRATION CERTIFICATE

Customer	M/S. Government Primary Health Centre,	Certificate No:	AMS/LAB/289/4914
Name &	Checkkanoorani, Theni Main Road,	Issuing Date:	30/09/2023
Address:	Madurai District-625514.	Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Micropipette	Id. No:	PHC/LAB/VMP/001
Make:	Thermo Scientific	Department:	Laboratory
Model:	Finnpipette'F3	Calibrated On:	30/09/2023
SI. No:	PW02827	Calibration Due	
Type:	Variable	Suggested:	29/09/2024
Range &	10 to 100μl	DUC Condition on	
Least Count:		Receipt:	OK

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electronic Weighing Scale	EWC/SF400C	PC/01/23/ATL/0512-02	31/01/2024
2	Thermo Hygrometer	Beetech/088	PC/01/23/ATL/0210-02	31/01/2024

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

Test ResultsVolume μl:

SL.	DUC Reading	Standard Reading (µl)		Correction
No.	(μl)	Trial 1	Trial 2	(μl)
1	10	10.0	10.0	+00.0
2	20	20.0	20.0	+00.0
3	50	50.0	50.0	+00.0
4	100	100.0	100.0	+00.0

Allowable Tolerance: User Defined

Note:

1. Calibration Points were selected as per customer request.

2. Each reading was taken by Repeatability Test(i.e.) each reading in a row is a mean of 10 reading.

3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:

(Calibration Engineer)

amshealthar mad@gmail.com

Approved by: (Technical Manager)

94434 74610, 93455 56148

aceagm@gmail.com