



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - AUGUST 2022



Lab Name Govt. PHC, Kadalaiyur
Constituent Group Chemistry III
PT item Lyophilized Serum

Lab No 12995
Date of Result Entered : 13/08/2022
Date of Report Published ; 05/09/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1833	104.52	7.14	7.46	110.7 mg/dL	0.83	0.35
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1780	6.46	8.21	0.53	7.61 mg/dL	2.17	0.03
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1717	115.04	5.56	6.39	121.7 mg/dL	1.04	0.31

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
Govt. PHC, Kadalaiyur
KADALAIYUR
Kadalaiyur
TAMILNADU628902

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in

Dey
MEDICAL OFFICER
GOVT. PRIMARY HEALTH CENTRE
KADALAIYUR

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

