



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - JULY 2022



Lab Name GOVT PRIMARY HEALTH CENTER KADALAIYUR  
Constituent Group Chemistry III  
PT item Lyophilized Serum

Lab No 12995  
Date of Result Entered : 18/07/2022  
Date of Report Published : 05/08/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1864	295.51	7.14	21.09	359.9 mg/dL	3.05	0.98
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1839	3.40	9.21	0.31	4.07 mg/dL	2.14	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1803	117.79	6.19	7.29	136.3 mg/dL	2.54	0.34

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

LAB ADDRESS :  
GOVT PRIMARY HEALTH CENTER KADALAIYUR  
KADALAIYUR  
KADALAIYUR  
TAMILNADU628902

*D.P.*  
MEDICAL OFFICER  
GOVT. PRIMARY HEALTH CENTRE  
KADALAIYUR

*Pamela Christudoss*

