

		<u>CALI</u>	<u>BRA</u>	<u>TION</u>	CERT	<u>IFICAT</u>	<u>E</u>		
ERTIFICATE NO: SBS/CL/23/13000			MEC		CHANICAL Pa		Page No:1 of 1		
ssue Date			06-10-2023						
SRF No & Date				SRF/23/00605-0005 & 05-10-2023					
Receipt Date				05-10-2023					
Calibration Date				05-10-2023					
Calibration Due				04-10-2024					
Customer Name & A	ddress								
GOVERNMENT PRIM	MARY HEALTH	CENTRE,							
CHENDUR-604302,V	ILLUPURAM D	ISTRICT.							
		De	tails of D	evice Un	der Calibra	tion (DUC)			
Description : CENTRIFUGE			Make & Model		:	B.D INSTRUMENTATION & BDI-152			
Range :	3500	3500 RPM		Sr. No		;	NA		
Resolution :		1 RPM		Identification No		:	NA		
UC Condition : Satisfactory			Location			LABORATORY			
	Env	ironmental (Condition	ns & Star	dard Oper	ating Proced	ure Details		
Environmental Details Temperature		e: 24.3 °	: 24.3 °C Relative H		midity 54 % Rh				
Calibration Procedure No SBS/CP/ML/0		/04	04 Calibration		done at ONSITE				
			Refer	rence Sta	ndards De	tails			
S.No Description	No Description Make/			SI No:		Certificate No		Validity	
1 Digital Tachometer Lutror			Lutron 8	n & DT-2238		JRPM-CCTR-A&S-2022-0036		03-10-2023	
CALIBRATION RESI	II TS					•			

CALIBRATION RESULTS

S.No	DEVICE UNDER CALIBRATION READINGS	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY (±)	
	RPM	RPM	RPM	%	
1	1000	999.9	0.1	4.2	
2	2000	1996	4	4.2	
3	3000	2997	3	4.2	

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- ${\small 6. \ Equipment\ used\ for\ Calibration\ were\ calibrated\ \&\ traceable\ to\ National\ \&\ International\ Standards.} \\$

Calibrated By

(Calibration Engineer)
M.DINESH



Quality Manager
(C.SIVABALAN)

Chief Executive

Authorised Signatory

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087, E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.