



CHRISTIAN MEDICAL COLLEGE
 DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - APRIL 2023



PC-1024

9267

20/04/2023

06/05/2023

Lab Name: Govt. PHC, Vellalankottal
 Consultant Group: Chemistry III
 PT Item: Lyophilized human serum based

Lab No: _____
 Date of Result Entered: _____
 Date of Report Published: _____

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	Participants			Your Value	SDI	U
					IV	CV	SD			
1	GLUCOSE III	GOO-POD III	Robonik (Semi automatic / Fully Automatic)	1832	171.45	6.04	7.21	126 mg/dL	0.91	0.34
2	UREA III	Others (Urease Berthelot / Nid Dye) III	Any Analyser (Automation / Semi Automation)	56	43.27	11.10	4.80	40 mg/dL	-0.58	1.28
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.18	8.09	0.18	2 mg/dL	-0.91	0.01
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	115 mg/dL	0.59	0.28

SDI Range		Interpretation
Within -1.00 to +1.00		Excellent.
Within ±1.01 to ±2.00		Good.
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.
Beyond ±3.0		Unacceptable performance. Action Signal.

LAB ADDRESS :
 Govt. PHC, Vellalankottal
 THE MEDICAL OFFICER, GOVT. PRIMARY HEALTH CENTRE, VELLALANKOTTAL, VELLALANKOTTAL-PO, KAYATHAR TK, THOOTHUKUDI .
 Vellalankottal
 TAMILNADU-628952

Coordinator Contact Details:
 Email: ccmc@cmcvellore.ac.in
 Contact Number: 0416-2223102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

S.No	Date	Name of the Test	Method	Name of the Lab	Batch No.	Value mg/dL	Report updated date	Report updated date
1	17-04-23	GLUCOSE	GOO-POD	CORAL	047200544/0024	126 mg/dL	17-04-23	0-91 SX
2	17-04-23	CHOLESTEROL	CHOD-PAP	LAB. CARE	CH0201026/02024	130 mg/dL	17-04-23	0-59 EX
3	17-04-23	CREATININE	Jaffes method	ERBA	50122436/9124	2.0 mg/dL	17-04-23	-0.91 SX
4	17-04-23	UREA	BERTHLOT	BIO LAB	09/may 24	40 mg/dL	17-04-23	-0.68 SX