



ALLWIN MEDICAL SYSTEMS

No. 2/4, 2nd Street, 1st Floor, Jayalakshmi Nagar,
Kattupakkam, Chennai - 600 056, Tamilnadu, India,

Cell : 9443663366

Email : allwinmedicalsyste.ms@gmail.com

CALIBRATION CERTIFICATE

Date: 29.09.2023

Calibration Name & Address		Certificate No	AWMS70C23
Government Primary Health Centre Natteri			
Details of Device Under Calibration (DUC)			
Description	: Micro Pipette	Make:	Thermo scientific
Range	: 10 - 100ul, 100 - 1000ul	Model :	F3
Least Count	: ± 2	Sr.No :	RW12599, RW10090
DUC Condition	:	Location :	Lab
Environmental Conditions & Calibration Procedure Details			
Environmental Details	Temperature: 25°C	Relative humidity	64%RH
Sample Calibration date	06.03.23	Calibration Done at	ON SITE

CALIBRATION RESULTS

S.No	DEVICE UNDER CALIBRATION READINGS	STANDARD INSTRUMENTS READINGS			DEVIATION	EXPANDED UNCERTAINTY(\pm)2ul
	ul					
1	100 - 1000ul	1000ul	1000ul	1000ul	Ok	Ok
2	10 - 100ul	100ul	100ul	100ul	Ok	Ok

Remarks

- 1.This Calibration Certificate Shall not be reproduced except in full, without written approval of the laboratory
- 2.The user Should be determine the suitability of the instrument for its intended use
- 3.The Recalibration interval should be determine on the User requirement.
- 4.The results Stated in this Certificate relate only to the item Calibrated
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for coverage factor K=2.00
- 6.Equipment Used for Calibration were Calibrated & Traceable to National & International Standards

Calibrated By

Authorized Signatory

