



CHRISTIAN MEDICAL COLLEGE
 DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - AUGUST 2023



Lab Name: Govt. PHC, Vadamalapuram
 Constituent Group: Chemistry III
 PT item: Lyophilized human serum based

Lab No: 13158
 Date of Result Entered: 19/08/2023
 Date of Report Published: 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	169.1 mg/dL	-0.72	0.42
2	UREA III	Others (Urease Berthelot / Ned Dye)	Any Analyser (Automation / Semi Automation)	59	28.92	13.99	4.05	29.5 mg/dL	0.14	1.05
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.8 mg/dL	0.34	0.01
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	116 mg/dL	0.47	0.22
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :
 Govt. PHC, Vadamalapuram
 CLINICAL LABORATORY, SIVAKASI
 Vadamalapuram
 TAMILNADU626130

Coordinator Contact Details:
 Email:clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

GOVT. PHC, VADAMALAPURAM - 626130
 VADAMALAPURAM - 626130