


CALIBRATION REPORT	
Report No : HBMS/CAL/23-24	Calibration Date : 03.10.2023
	Calibration valid : 02.04.2024
1.1.CUSTOMER DETAILS:	
Name and Address of the organization	M/s, URBAN PRIMARY HEALTH CENTRE, DADAGAPATTI, SALEM.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	MICRO PIPETTE
i Manufactured by	THERMOSCIENTIFIC
ii Model	FINNPIPETTE F3
iii Value	10 µl -100 µl
iv Serial No	RW09094

Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Calibration test Report.

Number of measurement	Channel 1 (Volume)		
	10µl	50µl	100µl
1	10.06	50.01	100.16
2	10.08	50.04	100.14
3	10.02	50.03	100.15
4	10.04	50.06	100.03
5	10.07	49.99	100.02
6	10.00	50.02	100.09



For Heart Beat Medical Systems,

(Authorized signatory)



MD Reg No: TN/CBE/MD42/00016

99400 28355, 9751206106

No.42/1, Navaindia Road, Peelamedu,

GSTIN: 33AAJFH1833M1ZB

88833 15662

CBE - 04, hbmedicalsystem@gmail.com