


CALIBRATION REPORT	
Report No : HBMS/CAL/23-24	Calibration Date : 03.10.2023 Calibration valid : 02.04.2024
1.1.CUSTOMER DETAILS:	
Name and Address of the organization	M/s, URBAN PRIMARY HEALTH CENTRE, DADAGAPATTI, SALEM.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	MICRO PIPETTE
i Manufactured by	P FACT CE
ii Model	----
iii Value	10 µl -100 µl
iv Serial No	235336


Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Calibration test Report.

Number of measurement	Channel 1 (Volume)		
	10µl	50µl	100µl
1	10.05	50.07	100.13
2	10.07	50.05	100.16
3	10.04	50.03	100.14
4	10.03	50.09	100.07
5	10.09	49.98	100.05
6	10.01	50.06	100.11

For Heart Beat Medical Systems,
(Authorized signatory) 



MD Reg No: TN/CBE/MD42/00016

99400 28355, 9751206106

No.42/1, Navaindia Road, Peelamedu,

GSTIN: 33AAJFH1833M1ZB

88833 15662

CBE – 04, hbmedicalsystem@gmail.com