

# HOPE ULTRA SYSTEMS

No: 2/35, MAIN ROAD, MAMSAPURAM (PO),  
SIVAKASI, VIRUDHUNAGAR-626124.  
E-mail : [hopeultrasystems@gmail.com](mailto:hopeultrasystems@gmail.com)

GSTIN. 33DLMPK3057D1Z9  
Mobile: +919487402761, +919715223192.

## CALIBRATION CERTIFICATE

Requested By

**GUPHC, SRIVILLIPUTHUR.**

### Scope:

|                             |                                      |
|-----------------------------|--------------------------------------|
| Equipment Name : CENTRIFUGE | Asset no : PHC/SVP/23                |
| Make : REMI                 | Calibration Done at : customer place |
| Model : C-852               | Calibration Date : 10/01/2023        |
| Sl no : Z CEN-13683         | Next call due on : 09/01/2024        |

| Sl. No. | TEST DESCRIPTION                     | RESULT |
|---------|--------------------------------------|--------|
| 01      | Power supply                         | Ok.    |
| 02      | Timer/Speed control/ Selector switch | Ok     |
| 03      | Door Alarms                          | Ok     |
| 04      | Indication lamp                      | Ok     |

Error Acceptance : Not Applicable (user defined)

Equipment Status : Working in Good condition

Materials used for testing and calibration:

1. Digital Multimeter-Mastech-MS830L
2. Taco meter



**MEDICAL OFFICER**  
Govt. Urban PHC  
Srivilliputhur - 626 125.

