

ICTC ANADONI - 200 TONN LIBERATOR POLY AMON
TEMPERATURE LOG FORM

Equipment Number: 18 Refrigerator Model & Year: 2017
Location: 200 TONN LIBERATOR POLY AMON
Temperature: 2.2°C

DATE	TEMPERATURE	REMARKS
11/11/2017	2.2	
12/11/2017	2.2	
13/11/2017	2.2	
14/11/2017	2.2	
15/11/2017	2.2	
16/11/2017	2.2	
17/11/2017	2.2	
18/11/2017	2.2	
19/11/2017	2.2	
20/11/2017	2.2	
21/11/2017	2.2	
22/11/2017	2.2	
23/11/2017	2.2	
24/11/2017	2.2	
25/11/2017	2.2	
26/11/2017	2.2	
27/11/2017	2.2	
28/11/2017	2.2	
29/11/2017	2.2	
30/11/2017	2.2	
01/12/2017	2.2	
02/12/2017	2.2	
03/12/2017	2.2	
04/12/2017	2.2	
05/12/2017	2.2	
06/12/2017	2.2	
07/12/2017	2.2	
08/12/2017	2.2	
09/12/2017	2.2	
10/12/2017	2.2	
11/12/2017	2.2	
12/12/2017	2.2	
13/12/2017	2.2	
14/12/2017	2.2	
15/12/2017	2.2	
16/12/2017	2.2	
17/12/2017	2.2	
18/12/2017	2.2	
19/12/2017	2.2	
20/12/2017	2.2	
21/12/2017	2.2	
22/12/2017	2.2	
23/12/2017	2.2	
24/12/2017	2.2	
25/12/2017	2.2	
26/12/2017	2.2	
27/12/2017	2.2	
28/12/2017	2.2	
29/12/2017	2.2	
30/12/2017	2.2	
31/12/2017	2.2	

Model No. GVR 225 AC E801583
ICE LINED REFRIGERATOR (LR)

IMPORTANT INFORMATION

- Do the internal service of the refrigerator at a safe temperature so that the contents are not damaged in the whole compartment.
- Do not use any compressed air to clean the internal parts (compressors, fan, coils, etc.) because the moisture is harmful to the refrigerator.
- The refrigerator is not to be used as a freezer. Do not change the temperature.
- Do not use the door as a handle or a support.

SET UP

USE/OPERATION

MAINTENANCE

TECHNICIAN SERVICE (6 MONTHS)

GVR 225 AC 3 EASY STEPS TO INSTALL

1. Check the installation site. The refrigerator should be installed on a level surface. The floor should be clean and dry. The refrigerator should be placed in a well-ventilated area. The refrigerator should be placed at least 5 cm away from the wall and other objects.
2. Level the refrigerator. Use a spirit level to check the level of the refrigerator. Adjust the leveling legs at the back of the refrigerator to make it level. The refrigerator should be level within 2 mm.
3. Plug the power cord. Plug the power cord into a grounded wall outlet. Do not use an extension cord. The refrigerator should be plugged into a dedicated circuit.

For more information, contact us on help@surechill.com or marketing@goody.com

SIMCO CALIBRATION LABORATORY

PCPC

SIMCO
AN ISO 9001 : 2015
CERTIFIED COMPANY

(A Division of : Sharp Industrial Machinery Maintenance Co. Pvt. Ltd.)

10-3-74/27 Plot No 151, Street No.3, Teacher's Colony, East Marredpally, Secunderabad - 26, Telangana
Tel: 040 2773 2341/342, 2773 1510, Mob: 77299 91231/33, 98480 46524.

E-mail: simco.hydr@gmail.com, Info@simco-india.com, www.simcocalibrationlaboratory.com



CALIBRATION CERTIFICATE

In accordance with ISO / IEC-17025 : 2017

CC-2806

F10-CC-03

Page : 1 of 2

Certificate No. : SL2304TL0287-005	Issue Date : 04-05-2023
1. Customer Name & Address: M/s.State Reference Laboratory Of HIV Testing (APSACS), Department Of Microbiology, Kurnool Medical College, Kurnool, 518 002.	ULR - CC 2 8 0 6 2 3 0 0 0 0 2 6 0 8 F Reference Date : 29-04-2023 Calibration Date : 30-04-2023 Calibration Due Date : 29-04-2024

2. Details of Unit Under Calibration:

Description : Digital Thermo Hygrometer With Probe	Model No. : HTC-2
Make : HTC	Id No. : APSACS/ICTC/AH/ADN/THM/01
Range : -50 to 70 °C / 10 to 99 %RH	
Resolution : 0.1 °C / 1 %RH	

3. Detail of Standard Instruments Used :

Instrument Used	SI / Id No	Valid up to	Certificate No.
Digital Thermo Hygrometer With Sensor	5180091/2027760 4	16-07-2023	30056478
Digital Temperature Indicator With SPRT Sensor	935-14-95 H	20-04-2024	TSC/22-23/18859-2

4. Environmental Conditions: Standard Temperature : (25±4) °C Relative Humidity : (50±20) % RH

5. Calibration Procedure: SOP-TL-01 / SOP-TL-04

6 Thermal Calibration: Temperature & Relative Humidity

7. Remarks:

- The instrument/equipment is in good condition and was calibrated at Lab.
- This certificate pertains only to the item calibrated.
- The calibration results reported in this certificate are valid at the time of and at the stated environmental conditions.
- The calibration interval is determined based on customer's requirements.
- The calibration is traceable to National standards as per traceability details given in the certificate.
- This calibration certificate shall not be reproduced in full, except with prior written approval of Managing Director, SIMCO Calibration Laboratory.
- This calibration certificate is meant for scientific and industrial purpose only.
- The NABL Symbol is used as per NABL guidelines in NABL-133.
- The Expanded Uncertainty is reported approximately at 95.45% confidence level with coverage factor $k=2$.

Shivaji
Calibrated by

P.A. Anandam
Mrs. P.A. Anandam
Technical Head
Authorised Signatory

SIMCO CALIBRATION LABORATORY

(A Division of : Sharp Industrial Machinery Maintenance Co. Pvt. Ltd.)

10-3-74/27, Plot No.151, Street No.3, Teacher's Colony, East Marredpally, Secunderabad - 26, Telangana
Tel: 040 2773 2341/342, 2773 1510, Mob: 77299 91231/33, 98480 46524.

E-mail: simcohyd@gmail.com, Info@simco-india.com, www.simcocalibrationlaboratory.com

CALIBRATION CERTIFICATE

In accordance with ISO / IEC-17025 : 2017

SIMCOTM
AN ISO 9001 : 2015
CERTIFIED COMPANY



CC-2806

F10-CC-03

Page : 2 of 2

Certificate No. : SL2304TL0287-005

ULR - CC 2 8 0 6 2 3 0 0 0 0 2 6 0 8 F

8. Calibration Results:

a. Temperature @ 50% RH

S No.	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Expanded Uncertainty in (±°C)
1	9.95	10.1	0.15	0.9
2	19.97	20.2	0.23	0.9
3	39.94	40.3	0.36	0.9
4	59.89	60.3	0.41	0.9


b. Relative Humidity @ 29°C

S No.	Standard Reading (%RH)	UUC Reading (%RH)	Error (%RH)	Expanded Uncertainty in (± %RH)
1	19.79	20	0.21	1.2
2	34.64	35	0.36	1.2
3	60.48	61	0.52	1.2
4	94.11	95	0.89	1.2

c. Temperature (Probe)

S No.	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Expanded Uncertainty in (±°C)
1	-30.032	-30.2	-0.168	0.4
2	0.038	0.1	0.062	0.4
3	10.026	10.1	0.074	0.4
4	40.044	40.2	0.156	0.4
5	60.011	60.3	0.289	0.4


Calibrated by


Mrs. P.A. Anandam
Technical Head
Authorised Signatory