



G MULTI SERVICES

78, Yahooob Street, Gobichettipalayam, Erode Dist, Tamil Nadu. Pin: 638 452.
 gmultiservicesgobi@gmail.com. Cell : 96597 20654.
 Bio Medical Instruments Services

Er. A. Ibathul Raziq DECE., BE (ICE)
 Bio Medical Engineer

Calibration Protocol & Report for Medical Equipment

Equipment : PIPETTE
 MAKE : FINNPIPEPETTE F3
 MODEL : 100-1000µl

PROTECT CLASS /TYPE : Risk Class II
 S. No. : SU/PI/002
 Location : GPHC, SUMAITHANGI.

S.No	Measurement Parameters	Status	Remarks
1	Electrical Test (a) Safety Standards (range : 210-240 V) (b) Equipotential ground & Earthing (c) Risk of shock/fire Hazards, if any (d) Chassis & Insulation Resistance (e) Earth & Chassis Leakage Current (f) Battery conditions & charging status	----	-
2	Physical Condition Test (a) Cabinet & Visible damages, if any (b) Risk of mechanical damages, if any	No No	Checked OK
3	Performance Test (e) Mean 99.3µl (f) A -0.5 µl (g) Acc % -0.71 (h) CV% -0.36	OK OK OK OK	
4	Calibration Test min volume 100 µl max Volume 1000 µl (e) A+or - 1.5, 5 µl (f) S 1.5 µl max Vol. 1000 µl (g) Acc+ or - 1.50, 0.50 (h) CV% 0.60, 0.10	OK OK OK OK	Tested
Formulated By: Edition Research Laboratory		Issued By:	
Endorsed By : Research team head		User Department:	
Reviewed On: 03.09.23 Next Review : 02.09.24			