

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. UPHC, Pernampattu
 Lab No 13066
 Month May
 Year 2023
 Constituent Group Chemistry III

Click on the analyte to view Graphical Data

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

[All Analyser Result](#) [Print](#)

Date of Result Entered : 12/05/2023

Date of Report Published : 05/06/2023

Sl. No.	Analyte	Method / Principle Name	Analyser Name	No. of Participants	CV	Participants		Your Value	SDI	SDI
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	202.3 mg/dL	2.28	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.8 mg/dL	0.30	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	110.4 mg/dL	-0.72	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Method	Submit Lab Results	View Lab Monthly Summary
View Summary	Change Password	Home	Sign-out