

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. UPHC, Pernampattu
 Lab No 13066
 Month April
 Year 2023
 Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered : 10/04/2023

Date of Report Published : 06/05/2023

SLNo.	Analyte	Reference/Procedure Name	Analyser Name	No of Participants	DV	Participants		Turn Value	SDI	U
						CV	CV			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	118.4 mg/dL	-0.15	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.2 mg/dL	0.23	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	115.2 mg/dL	0.62	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within +1.01 to +2.00	Good.
Within +2.01 to +2.99	Accept with caution. Warning Signal.
Beyond +3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Control Lab Results	View Lab Monthly Summary
Update Summary	Change Password	Home	Signout