VIEW LAB MONTHLY SUMMARY

| | |
|---|---|
| Lab Name | Govt. UPHC, Pernampattu |
| Lab No | 13066 |
| Month | June |
| Year | 2023 |
| Constituent Group | Chemistry III |

Click on the analyte to view Graphical Data

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

All Analyser Result | Print

Date of Result Entered : 13/06/2023

Date of Report Published : 04/07/2023

| SI.No | Analyte | Group / Principle Name | Analyser Name | No of Participants | [illegible] | [illegible] | [illegible] | Year Value | SDI | [illegible] |
|---|---|---|---|---|---|---|---|---|---|---|
| 1 | GLUCOSE III | GOD-POD III | Robonik ( Semi automatic / Fully Automatic ) | 1823 | 286.63 | 5.17 | 14.92 | 308.3 mg/dL | 1.46 | 0.69 |
| 2 | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik ( Semi automatic / Fully Automatic ) | 1848 | 2.57 | 6.64 | 0.17 | 2.6 mg/dL | 0.18 | 0.01 |
| 3 | CHOLESTROL III | CHOD-PAP III | Robonik ( Semi automatic / Fully Automatic ) | 1804 | 113.73 | 4.46 | 5.07 | 108.5 mg/dL | -1.03 | 0.24 |

| SDI Range | Interpretation |
|---|---|
| Within -1.00 to +1.00 | Excellent. |
| Within ±1.01 to ±2.00 | Good. |
| Within ±2.01 to ±2.99 | Accept with caution. Warning Signal. |
| Beyond ±3.0 | Unacceptable performance. Action Signal. |

Links

| | | | |
|---|---|---|---|
| Modify Address | Modify Methods | Submit Lab Results | View Lab Monthly Summary |
| Yearly Summary | Change Password | Home | Logout |