

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE
SCHEME



EQAS YEARLY SUMMARY - 2022

Lab Name Govt. PHC, Murungapatti Lab No 12628

Constituent Group Chemistry III

PT item Lyophilized human serum based Date of Report Published 20-02-2023

Group I : Laboratories who reported 10 to 12 months during 2022 Group II : Laboratories who reported 6 to 9 months during 2022

Group III: Laboratories who reported less than 6 months during 2022

SI.No	Parameter	Rank Group	No of Labs	OMSDI	Grade
1	GLUCOSE III	Group I	1941	0.89	A
2	CREATININE III	Group I	1927	0.80	Α
3	CHOLESTROL III	Group I	1923	0.65	Α

Only first 2 Groups are Graded

OMSDI Range	Grade	Remarks	
0.00 - 1.00	Α	Excellent	
1.01 - 2.00	В	Good	
2.01 - 2.99	С	Satisfactory	
Beyond ±3.00	D	Unacceptable	

LAB ADDRESS

Govt. PHC, Murungapatti Gov primary health center murungapatty Murungapatti TAMILNADU636307

Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact number:0416-2283102 Panela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College,

Vellore

Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
******** End of Report *******