

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. UPHC, Suriyampalayam

Details About Robust Analysis

Lab No 10706

Detail About Monthly Summary

Month September

Detail about SDI

Year 2023

Constituent Group Chemistry III

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 20/09/2023

Date of Report Published : 05/10/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	36	312.41	4.03	12.59	320 mg/dL	0.60	4.20
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	38	1.30	7.23	0.09	1.2 mg/dL	-1.06	0.03
3	CHOLESTROL III	CHOD-PAP III	Biosystems	35	113.55	3.41	3.87	110 mg/dL	-0.92	1.31

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out