

CALIBRATION CERTIFICATE

Date: 27.09.2023

<u>Calibration Name & Address</u>		<u>Certificate No</u>		AWMS29C23	
Government Primary Health Centre Onnupuram					
Details of Device Under Calibration (DUC)					
Description : Micro Pipette		Make: Micromax			
Range : 10 - 50ul, 100 - 1000ul		Model :			
Least Count : ± 2		Sr.No :			
DUC Condition :		Location : Lab			
Environmental Conditions & Calibration Procedure Details					
<u>Environmental Details</u>		Temperature: 25°C		Relative humidity	
				49%RH	
<u>Sample Calibration date</u>		06.03.23		<u>Calibration Done at</u>	
				ON SITE	

CALIBRATION RESULTS

S.No	DEVICE UNDER CALIBRATION READINGS	STANDARD INSTRUMENTS READINGS			DEVIATION	EXPANDED UNCERTAINTY(\pm)2ul
ul						
1	100 - 1000ul	1000ul	1000ul	998ul	-2ul	Ok
2	10 - 50ul	50ul	50ul	50ul	Ok	Ok

Remarks

- 1.This Calibration Certificate Shall not be reproduced except in full, without written approval of the laboratory
- 2.The user Should be determine the suitability of the instrument for its intended use
- 3.The Recalibration interval should be determine on the User requirement.
- 4.The results Stated in this Certificate relate only to the item Calibrated
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for coverage factor K=2.00
- 6.Equipment Used for Calibration were Calibrated & Traceable to National & International Standards

Calibrated By



Authorized Signatory


