

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. UPHC, Suriyampalayam	Details About Robust Analysis
Lab No	10706	Detail About Monthly Summary
Month	August	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 19/08/2023

Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	34	172.03	5.51	9.48	170 mg/dL	-0.21	3.25
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	35	1.76	6.47	0.11	1.6 mg/dL	-1.40	0.04
3	CHOLESTROL III	CHOD-PAP III	Biosystems	33	114.98	3.90	4.49	120 mg/dL	1.12	1.56

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out