



Vanavil Calibrations
Make It Simple, But Significant!

CALIBRATION CERTIFICATE

Page no: 1 of 1

Customer Name & Address: M/S GOVERNMENT PRIMARY HEALTH CENTRE, THOLASAMPATTI - 636 503.	Certificate No:	VCPL/CC/23-24/36-004
	Date of Calibration:	05-10-2023
	Date of issue:	06-10-2023
	Date of Receipt:	05-10-2023
	Next call Date:	04-10-2024

Details of Device Under Calibration

Description:	BEMI AUTO ANALYZER	Make:	ROBONIK
Range:	-	Model/Type:	PRIESTEST TOUCH
Resolution:	-	Condition of DUC	GOOD
Serial Number	AT0960413RBK	Location:	LABORATORY
ID number	TGPHC/SAA/01	Cal. At:	ONSITE
Environmental Conditions		Calibration Method Used	
Temperature	25.2 °C	National/International standard:	IEC60601-1
Humidity	49 % Rh	Cal procedure no:	-

Master Details

SL no:	Description	ID No / SL No	Certificate No	Make/Model	Valid Till
1	Electrical Safety Analyzer	6014025	ACCS230581	FLUKE/ESA-612	9-Mar-24

Electrical Safety Report

S.NO	PARAMETERS	MEASURED VALUES μ A	ALLOWABLE LIMITS μ A
1	Earth Leakage	122.3	~5000 μ A For B,BF,CF
		MEASURED VALUES μ A	ALLOWABLE LIMITS μ A
2	Enclosure Leakage	187.5	~5000 μ A For B,BF,CF

Remarks:

1. DUC is Defined as Device Under Calibration.
2. Calibration Frequency decided by the customer.
3. The above instrument is working in good condition.

Calibrated By: _____

R.GOPALAKRISHNAN
 (CALIBRATION ENGINEER)

Authorised by: _____

T. Sudhakar
T. SUDHAKAR
 Technical Manager
 Vanavil Calibrations Pvt Ltd.

*** End of the Report ***

Vanavil Calibrations Pvt. Ltd

(A Unit of Rainbow Bio Medical Engineers)

S2, Rishi Mahara, Apartment, Indra Nagar, 250 Reddyar Main Road, Narasimhapatty, Salem - 636 004.

No. 96,97,98, Anandhamaya Apartments, 2nd Floor, KVR Flats, 2nd Cross Street, Thangam Nagar, Pallikarasa, Chennai - 600 100.

Call: 98545 13872 / 99945 23872 Ph: 0427-7565376

vanavilcalibrations@gmail.com
 www.rainbowbiomedicalengineers.com | www.vanavilcalibrations.com