

				CALIE	BRAT	TION	CERTI	FICATE			
CERTI	FICATE NO	:	SBS/CL/23	3/13082			ME	CHANICAL	Pa	age No:1 of 1	
Issue D	ate					07-10-20)23				
SRF No & Date				SRF/23/00668-0001 & 06-10-2023							
Receipt Date				06-10-2023							
Calibra	tion Date					06-10-2023					
Calibration Due				05-10-2024							
Custor	ner Name 8	& Addr	ess								
GOVE	RNMENT PI	RIMAR	Y HEALTH	CENTRE,							
SIKKA	L-623528,R	AMANA	ATHAPURAI	M DISTRICT							
				Det	ails of D	evice Un	der Calibrati	ion (DUC)			
Descri	otion	:	CENTRIFL	JGE		Make &	Model	:	B.D INSTRUMENT	ATION & BDI-152	
Range		:	3500	RPM		Sr. No			NA		
Resolution :		1 RPM			Identification No		:	NA			
DUC C	ondition	:	Satisfactor	у		Location	İ	::	LABORATORY		
			Envi	ronmental C	ondition	ns & Stai	ndard Opera	ting Procedu	re Details		
Environmental Details Temperature			: 25.4°C Relative Hu		midity	54% Rh					
Calibration Procedure No SBS/CP/ML/0			04 Calibration		lone at	ONSITE					
					Refe	rence Sta	andards Deta	ails			
S.No	S.No Description			Make/ SI No:		Certificate No		Validity			
1	Digital Tachometer			LINE SEIKI / 175-0034V		JRPM-CCTR-A&S-2023-0013		09-06-2024			

CALIBRATION RESULTS

S.No	DEVICE UNDER CALIBRATION	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY	
	RPM	RPM	RPM	%	
1	1000	999.1	0.9	4.2	
2	2000	1998.6	1.4	4.2	
3	3000	2999.5	0.5	4.2	

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- $2. \\ The user should determine the suitability of the instrument for its intended use.$
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

V S (Calibration Engineer)

K.SATHYAMOORTHY



Authorised Signatory

Quality Manager

Chief Executive

(C.SIVABALAN)

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087, E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.