

## VIEW LAB MONTHLY SUMMARY

<b>Lab Name</b>	<b>Govt. UPHC, Bykara</b>	<a href="#">Details About Robust Analysis</a>									
<b>Lab No</b>	<b>10775</b>	<a href="#">Detail About Monthly Summary</a>									
<b>Month</b>	<b>July</b>	<a href="#">Detail about SDI</a>									
<b>Year</b>	<b>2023</b>										
<b>Constituent Group</b>	<b>Chemistry III</b>										
		<a href="#">Click on the analyte to view Graphical Data</a>				<a href="#">All Analyser Result</a>		<a href="#">Print</a>			
		Date of Result Entered : 17/07/2023									
		Date of Report Published : 07/08/2023									
SI.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U	
						CV	SD				
1	<a href="#">GLUCOSE III</a>	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	96 mg/dL	-2.20	0.28	
2	<a href="#">CREATININE III</a>	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1850	5.83	6.00	0.35	6.3 mg/dL	1.34	0.02	
3	<a href="#">CHOLESTROL III</a>	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.66	113 mg/dL	0.00	0.22	
SDI Range			Interpretation								
Within -1.00 to +1.00			Excellent.								
Within ±1.01 to ±2.00			Good.								
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.								
Beyond ±3.0			Unacceptable performance. Action Signal.								

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>