



CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/13382		MEDICAL DEVICES		Page No:1 of 1
Issue Date		07-10-2023		
SRF No & Date		SRF/23/00736-0001 & 07-10-2023		
Receipt Date		07-10-2023		
Calibration Date		07-10-2023		
Calibration Due		06-10-2024		
Customer Name & Address				
GOVERNMENT PRIMARY HEALTH CENTRE, KARUPPAMPALAYAM-639003,KARUR DISTRICT.				
Details of Device Under Calibration (DUC)				
Description : SEMI AUTO ANALYZER		Make & Model : ROBONIK & PRIETEST TOUCH		
Range : MULTI		Sr. No : AT2230211RBK		
Resolution : MULTI		Identification No : NA		
DUC Condition : SATISFACTORY		Location : LABORATORY		
Environmental Conditions & Calibration Procedure Details				
Environmental Details		Temperature:25.6 ° C		Relative Humidity
Calibration Procedure No		SBS/CP/MD/20		52% RH
		Calibration done at		ONSITE
Reference Standards Details				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

ELECTRICAL SAFETY

RESULTS

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resistance >20MΩ	Measured values in MΩ	Uncertainty in % (±)
		88	13.92
2	Earth Leakage <5000µAfor B,BF,CF	Measured values in µA	Uncertainty in % (±)
		118	8.1
3	Enclosure Leakage <500µAfor B,BF,CF	Measured values in µA	Uncertainty in % (±)
		228	7.2

REMARKS

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

K.Sathyamoorthy
(Calibration Engineer)
K.SATHYAMOORTHY



Authorised Signatory

C.Shanmugaraaj
Technical Manager
C.SHANMUGARAJ

Chief Executive

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.