



CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/13023

Page. No : 1 of 1

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE, DEVIPATTINAM, RAMANATHAPURAM DISTRICT.	SRF No.	SRF/23/00655-0001
	SRF Date	06-10-2023
	Date of Receipt	05-10-2023
	Date of Calibration	06-10-2023
	Due Date for Calibration	05-10-2024
	Issue Date	07-10-2023

Details of Unit Under Calibration

Description	Micro Pipette	Make	THERMO SCIENTIFIC
Range	100-1000µl	Model	FINNPIPETTE F3
Resolution	1µl	Material	PVC
Serial Number	RW13746	Operating Range	100-1000µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition				Calibration Method Used	
Temperature (°C)	23.9	Humidity (%RH)	55	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSPL22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319


Result of Calibration in µl										
Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
1	100	99.92	99.91	99.92	99.91	99.90	99.91	-0.09	0.01	0.47
		99.92	99.91	99.91	99.92	99.92				
2	500	499.67	499.68	499.67	499.68	499.67	499.68	-0.32	0.01	0.47
		499.68	499.67	499.68	499.68	499.67				
3	1000	999.66	999.65	999.64	999.63	999.62	999.63	-0.37	0.02	0.47
		999.65	999.64	999.62	999.62	999.61				

Remarks

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
2. The user should determine the suitability of the instrument for its intended use.
3. The recalibration interval should be determined on the user requirement.
4. The results stated in this certificate relate only to the item calibrated.
5. Equipment used for Calibration were calibrated & traceable to National & International Standards
6. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
7. Calibration Liquid Used: Distilled or Deionized water, certified, Grade3 as specified in ISO 3696.

Calibrated By,

(Calibration Engineer)
 M.RAGUL



Authorised by:

Quality Manager/Chief Executive
 C.SIVABALAN

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
 E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.

CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/13024

Page No : 1 of 1

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE, DEVIPATTINAM,RAMANATHAPURAM DISTRICT.	SRF No.	SRF/23/00655-0002
	SRF Date	06-10-2023
	Date of Receipt	05-10-2023
	Date of Calibration	06-10-2023
	Due Date for Calibration	05-10-2024
	Issue Date	07-10-2023

Details of Unit Under Calibration

Description	Micro Pipette	Make	THERMO SCIENTIFIC
Range	10-100µl	Model	FINNPIPETTE F3
Resolution	0.2µl	Material	PVC
Serial Number	RW03382	Operating Range	10-100µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition				Calibration Method Used	
Temperature (°C)	23.7	Humidity (%RH)	53	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSP22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319

Result of Calibration in µl											
Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)	
1	10.0	9.87	9.86	9.87	9.86	9.87	9.86	-0.14	0.01	0.47	
		9.86	9.82	9.86	9.87	9.86					
2	50.0	49.46	49.47	49.48	49.47	49.46	49.48	-0.52	0.01	0.47	
		49.47	49.48	49.48	49.49	49.49					
3	100.0	99.37	99.38	99.39	99.31	99.40	99.40	-0.60	0.04	0.47	
		99.42	99.42	99.44	99.44	99.45					

Remarks

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- Equipment used for Calibration were calibrated & traceable to National & International Standards
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
- Calibration Liquid Used: Distilled or Deionized water conforming Grade 3 as specified in ISO 3696.

Calibrated By,


 (Calibration Engineer)
 M.RAGUL



Authorised by:


 (Quality Manager/Chief Executive)
 C.SIVABALAN

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.



CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/13025

Page. No: 1 of 1

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE, DEVIPATTINAM, RAMANATHAPURAM DISTRICT.	SRF No.	SRF/23/00655-0003
	SRF Date	06-10-2023
	Date of Receipt	05-10-2023
	Date of Calibration	06-10-2023
	Due Date for Calibration	05-10-2024
	Issue Date	07-10-2023

Details of Unit Under Calibration

Description	Micro Pipette	Make	THERMO SCIENTIFIC
Range	10-100µl	Model	FINNPIPETTE F3
Resolution	0.2µl	Material	PVC
Serial Number	RW10399	Operating Range	10-100µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition

Calibration Method Used

Temperature (°C)	23.5	Humidity (%RH)	51	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSP/22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319

Result of Calibration in µl


Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
1	10.0	9.94	9.93	9.94	9.92	9.94	9.93	-0.07	0.01	0.47
		9.92	9.92	9.94	9.94	9.94				
2	50.0	49.87	49.88	49.87	49.88	49.87	49.87	-0.13	0.01	0.47
		49.88	49.86	49.87	49.86	49.87				
3	100.0	99.86	99.87	99.88	99.87	99.88	99.87	-0.13	0.01	0.47
		99.87	99.86	99.86	99.88	99.86				

Remarks

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
2. The user should determine the suitability of the instrument for its intended use.
3. The recalibration interval should be determined on the user requirement.
4. The results stated in this certificate relate only to the item calibrated.
5. Equipment used for Calibration were calibrated & traceable to National & International Standards
6. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
7. Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,

(Calibration Engineer)
 M.RAGUL



Authorised by:

(Quality Manager/Chief Executive)
 C.SIVABALAN

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
 E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.

CALIBRATION CERTIFICATE

CERTIFICATE NO:	SBS/CL/23/13026	MECHANICAL	Page No:1 of 1	
Issue Date	07-10-2023			
SRF No & Date	SRF/23/00655-0004 & 06-10-2023			
Receipt Date	06-10-2023			
Calibration Date	06-10-2023			
Calibration Due	05-10-2024			
Customer Name & Address				
GOVERNMENT PRIMARY HEALTH CENTRE, DEVIPATTINAM,RAMANATHAPURAM DISTRICT.				
Details of Device Under Calibration (DUC)				
Description	CENTRIFUGE	Make & Model	REMI & C-852	
Range	3500 RPM	Sr. No	ZIFN25482	
Resolution	1 RPM	Identification No	NA	
DUC Condition	Satisfactory	Location	LABORATORY	
Environmental Conditions & Standard Operating Procedure Details				
Environmental Details	Temperature: 25.4°C	Relative Humidity	54% Rh	
Calibration Procedure No	SBS/CP/ML/04	Calibration done at	ONSITE	
Reference Standards Details				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Digital Tachometer	LINE SEIKI / 175-0034V	JRPM-CCTR-A&S-2023-0013	09-06-2024

CALIBRATION RESULTS

S.No	DEVICE UNDER CALIBRATION	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY
	RPM	RPM	RPM	%
1	1000	998.6	1.4	4.2
2	2000	1998.6	1.4	4.2
3	3000	2998.7	1.3	4.2

REMARKS


- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor $k=2.00$.
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By


(Calibration Engineer)
E.ESWAR



Authorised Signatory


Quality Manager Chief Executive
(C.SIVABALAN)

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.