



VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Erasakkanaicknoor	<a href="#">Details About Robust Analysis</a>
Lab No	12704	<a href="#">Detail About Monthly Summary</a>
Month	April	<a href="#">Detail about SDI</a>
Year	2023	
Constituent Group	Chemistry III	

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 20/04/2023

Date of Report Published : 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1832	119.45	6.04	7.21	119 mg/dL	-0.06	0.34
2	UREA III	Urease UV / GLDH III	Robonik ( Semi automatic / Fully Automatic )	169	42.32	10.68	4.52	42 mg/dL	-0.07	0.70
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1820	2.16	8.09	0.18	2.14 mg/dL	-0.11	0.01
4	URIC ACID III	Enzymatic / Uricase Colorimetric III	Robonik ( Semi automatic / Fully Automatic )	38	4.58	9.59	0.44	4.9 mg/dL	0.73	0.14
5	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1750	111.61	5.19	5.79	114 mg/dL	0.41	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out