



NEW LAB MONTHLY SUMMARY

Lab Name	Govt. Block PHC, R. Reddiapatti	<a href="#">Details About Robust Analysis</a>
Lab No	3929	<a href="#">Detail About Monthly Summary</a>
Month	August	<a href="#">Detail about SDI</a>
Year	2023	
Constituent Group	Chemistry III	

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered : 19/08/2023

Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI
						CV	SD		
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation)	64	174.34	5.46	9.51	175.5 mg/dL	0.12
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	1.7 mg/dL	-0.50
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	149	113.96	4.36	4.97	117 mg/dL	0.61

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
beyond ±3.0	Unacceptable performance. Action Signal.

Home	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Monthly Summary	Change Password	Home	Sign-out