

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. UPHC, Reddiyur
 Lab No 10832
 Month April
 Year 2023
 Constituent Group Chemistry III

[Details About Robust Analysis](#)[Detail About Monthly Summary](#)[Detail about SDI](#)[Click on the analyte to view Graphical Data](#)[All Analyser Result](#)[Print](#)

Date of Result Entered : 20/04/2023

Date of Report Published : 06/05/2023

Sl. No	Analyte	Method/Reagent/Unit	Analyser Name	No. of Participants	CV	Mean	SD	Target	Delta	Delta 2
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	120.1 mg/dL	0.09	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.3 mg/dL	0.80	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	110.9 mg/dL	-0.12	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.


 MEDICAL OFFICER,
 UPHC, REDDIYUR