

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. UPHC, Reddiyur

[Details About Robust Analysis](#)

Lab No 10832

[Detail About Monthly Summary](#)

Month June

[Detail about SDI](#)

Year 2023

Constituent Group Chemistry III

[Click on the analyte to view Graphical Data](#)[All Analyser Result](#)[Print](#)

Date of Result Entered : 20/06/2023

Date of Report Published : 04/07/2023

Sl. No.	Method / Constituent Name	Analyzer Name	Run ID	Run Date	Run Time	Run Status	Run Result	Run Unit	Run Error	Run SDI
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	298 mg/dL	0.09	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.8 mg/dL	1.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	101 mg/dL	-2.51	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

[Signature]
 MEDICAL OFFICER,
 UPHC, REDDIYUR

Links