

VIEW LAB MONTHLY SUMMARY

Lab Name **Govt. UPHC, Reddiyur**

Lab No **10832**

Month **July**

Year **2023**

Constituent Group **Chemistry III**

[Details About Robust Analysis](#)[Detail About Monthly Summary](#)[Detail about SDI](#)[Click on the analyte to view Graphical Data](#)[All Analyser Result](#)[Print](#)

Date of Result Entered : 20/07/2023

Date of Report Published : 07/08/2023

Sl. No	Analyte	Method	Instrument	1st Run	2nd Run	3rd Run	4th Run	5th Run	6th Run	7th Run	8th Run	9th Run	10th Run
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	102 mg/dL	-1.21	0.28			
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	6.3 mg/dL	1.34	0.32			
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	114 mg/dL	0.21	0.22			

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within +2.01 to +2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Abh
 MEDICAL OFFICER,
 UPHC, REDDIYUR

Links