

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. UPHC, Reddiyur

[Details About Robust Analysis](#)

Lab No 10832

[Detail About Monthly Summary](#)

Month August

[Detail about SDI](#)

Year 2023

Constituent Group Chemistry III

[Click on the analyte to view Graphical Data](#)[All Analyser Result](#)[Print](#)

Date of Result Entered : 20/08/2023

Date of Report Published : 01/09/2023

	Analyte	Method / Reagent / Principle	Analyzer Name	No. of Participants	CV	Participant	Participant	Participant	Participant	Participant	Participant
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	179 mg/dL	0.38	0.42	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.9 mg/dL	1.18	0.01	
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	110 mg/dL	-0.82	0.22	

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

[Signature]
 MEDICAL OFFICER,
 UPHC, REDDIYUR

Links