

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. UPHC, Reddiyur
 Lab No 10832
 Month September
 Year 2023
 Constituent Group Chemistry III

[Details About Robust Analysis](#)[Detail About Monthly Summary](#)[Detail about SDI](#)[Click on the analyte to view Graphical Data](#)[All Analyser Result](#)[Print](#)

Date of Result Entered : 19/09/2023

Date of Report Published : 05/10/2023

1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1870	310.86	4.82	14.98	313 mg/dL	0.14	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.29	8.14	0.10	1.3 mg/dL	0.10	0.00
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1783	113.62	3.80	4.32	112 mg/dL	-0.38	0.20

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

File
 MEDICAL OFFICER,
 UPHC, REDDIYUR