

			CAL	IBRAT	ION	CERT	<u>IFICA</u>	TE	
					MEDICAL DEVICES			Page No:1 of 1	
CERTIFICATE NO: SBS/CL/23/13144				06-10-2023					
SRF No & Date				SRF/23/00682-0001 & 05-10-2023 05-10-2023					
Receipt Date Calibration Date				05-10-2023 04-10-2024					
Cueto	ation Due omer Name & Ad	dress							
COM	RNMENT PRIMAYAPATTINAM-62	ARY HEAL	LTH CENTRE, MANATHAPUR	IAM DISTRI	CT.		· · · · · · · · · · · · · · · · · · ·		
				Details of D	evice	Jnder Calibra	tion (DUC)	
Descri	escription : SEMI AUTO ANALYZE			Sr. No			:	ROBONIK & PRIET AT1151010RBK	rest touch
Resol	ution	MULTI	CACTORY		Locat	ication No on	:	NA LABORATORY	
DUC	Condition	SATIO	Enviror	mental Con	dition	& Calibration	n Procedu	ire Details	
Environmental Details Temperature:				25.6°C		Relative Humidity Calibration done at		52% RH ONSITE	
Calibration Procedure No SBS/CP/MD			SBS/CP/MD/2	Reference Standards Details					
	[05-1-4C			Make/ SI No			Certificate No		Validity
S.No Description 1 Electrical Safety Analyser		Rigel Medical & 44L-1059		M-230809-16-4		10-08-20			

ELECTRICAL SAFETY				
MEASURED VALUES	EXPANDED UNCERTAINTY (±)			
Measured values in MΩ	Uncertainty in % (±)			
80	13.92			
Measured values in µA	Uncertainty in % (±)			
148	7.8			

RESULTS

S.no

1

2

3

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.

SPECIFICATION

Insulation Resisitance

Earth Leakage

<5000µAfor B,BF,CF

Enclosure Leakage

<500µAfor B,BF,CF

5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00

Measured values in µA

202

6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

Authorised Signatory





Technical	→ Manager
 C.SHANN	MUGARAJ

Chief	Executive
Cino	CAGOGIA

Uncertainty in % (±

7.3

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087, E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.