

LAB MONTHLY SUMMARY

Name: Govt. PHC, Sengal
No: 12297
Month: August
Year: 2023
Constituent Group: Chemistry III

[Details About Robust Analysis](#)
[Detail About Monthly Summary](#)
[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 19/08/2023
 Date of Report Published : 01/09/2023

No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	145 mg/dL	-3.4	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.27	0.12	1.63 mg/dL	-1.09	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.00	4.07	4.64	116 mg/dL	0.47	0.22

SDI Range	Interpretation
within -1.00 to -1.00	Excellent
within =1.01 to =2.00	Good
within =2.01 to =2.99	Accept with caution. Warning Signal
beyond =3.0	Unacceptable performance. Action Signal