

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Thimmaravuthampatti
 Lab No 12426
 Month June
 Year 2023
 Constituent Group Chemistry III

[Details About Robust Analysis](#)[Detail About Monthly Summary](#)[Detail about SDI](#)[Click on the analyte to view Graphical Data](#)**All Analyser Result****Print**

Date of Result Entered : 20/06/2023


Date of Report Published : 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	J
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	285.1 mg/dL	-0.10	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.85 mg/dL	1.64	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	114.6 mg/dL	0.17	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Facility Summary	Change Password	Home	Sign-out


 MEDICAL OFFICER
 Govt Primary Health Centre
 THIMMARAVUTHAMPATTI
 Duchergode Tk, Namakkal
 PIN 627 40