



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - JUNE 2023

June-2023  
PHC PCMC value



Lab Name	Govt. PHC, Padagam	Lab No	12922
Constituent Group	Chemistry III	Date of Result Entered :	20/06/2023
PT Item	Lyophilized human serum based	Date of Report Published :	04/07/2023

SLNO	Analyte	Method / Principle Name	Analyzer Name	No of Participant	CV	Participants		Your Value	SDI	U
						G	B			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	228.3 mg/dL	-3.94	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	118.22 mg/dL	680.29	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07	2.41 mg/dL	-21.96	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
$\geq 2.01$ to $\geq 2.99$	Accept with caution. Warning Signal.
Beyond $\geq 3.0$	Unacceptable performance. Action Signal.

LAB ADDRESS :  
Govt. PHC, Padagam  
GOVT PRIMARY HEALTH CENTER  
Padagam  
TAMILNADU 606904

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*

*[Signature]*  
MEDICAL OFFICER,  
Govt. Primary Health Centre,  
PADAGAM - 606 904.