

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. UPHC, Mayiladuthurai

Lab No 11310

Month August

Year 2023

Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[All Analyser Result](#)

[Print](#)

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 19/08/2023

Date of Report Published : 01/09/2023

Analyte	Method / Principle	Analysed Name	No of Samples	DV	Participants CV	Your Value	SDI	U	
1	GLUCOSE III	CCD-POD III	64	174.34	5.46	9.51	180 mg/dL	0.60	2.38
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	62	1.76	7.18	0.13	1.8 mg/dL	0.32	0.03
3	CHOLESTROL III	CHOD-PAP III	149	113.96	4.36	4.97	114 mg/dL	0.01	0.81
SDI Range			Interpretation						
Within -1.00 to +1.00			Excellent.						
Within ±1.01 to ±2.00			Good.						
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.						
Beyond ±3.0			Unacceptable performance. Action Signal.						

[Links](#)

Medical Officer
Urban Primary Health Centre,
Mayiladuthurai