

CALIBRATION CERTIFICATE

Page no: 1 of 1

Customer Name & Address: M/S GOVERNMENT PRIMARY HEALTH CENTRE, KARUPPUR - 636 012.		Certificate No:	VCPL/CC/23-24/39-004
		Date of Calibration:	05-10-2023
		Date of issue:	06-10-2023
		Date of Receipt:	05-10-2023
		Next cal Date:	04-10-2024
Details of Device Under Calibration			
Description:	SEMI AUTO ANALYZER	Make:	ROBONIK
Range:	-	Model/Type:	PRIETEST TOUCH
Resolution:	-	Condition of DUC	GOOD
Serial Number	AT1990509RBK	Location:	LABORATORY
ID number	KGPHC/SAAV01	Cal. At:	ONSITE
Environmental Conditions		Calibration Method Used	
Temperature	25.2 °C	National/International standard:	IEC60601-1
Humidity	49 % Rh	Cal procedure no:	-

Master Details

SL no:	Description	ID No / SL No	Certificate No	Make/Model	Valid Till
1	Electrical Safety Analyzer	6014025	ACCS230581	FLUKE/ESA-612	9-Mar-24

Electrical Safety Report

S.NO	PARAMETERS	MEASURED VALUES μ A	ALLOWABLE LIMITS μ A
1	Earth Leakage	122.3	<5000 μ A For B,BF,CF
		MEASURED VALUES μA	ALLOWABLE LIMITS μA
2	Enclosure Leakage	159.6	<5000 μ A For B,BF,CF

Remarks:

1. DUC is Defined as Device Under Calibration.
2. Calibration Frequency decided by the customer.
3. The above instrument is working in good condition.

Calibrated By:



R. GOPALAKRISHNAN
(CALIBRATION ENGINEER)



Authorised by:



T. SUDHAHAR,
Technical Manager

Vanavil Calibrations Pvt Ltd.

Vanavil Calibrations Pvt. Ltd

(A Unit of Rainbow Bio Medical Engineers)

52, Rishi Maharaj Apartment, Indra Nagar, 250 Reddiyur Main Road, Narasothipatty, Salem - 636 004.

97,98, Anadhamaya Apartments, 2nd Floor, KVR Flats, 2nd Cross Street, Thangam Nagar, Pallikaranai, Chennai - 600 100

Cell: 99949 13872 / 99949 23872 Ph: 0427-7965376

✉ vanavilcalibrations@gmail.com 🌐 www.rainbowbiomedicalengineers.com | www.vanavilcalibrations.com