



# Vanavil Calibrations

Make It Simple, But Significant !

## CALIBRATION CERTIFICATE

Page no: 1 of 1

<b>Customer Name &amp; Address:</b> M/S GOVERNMENT PRIMARY HEALTH CENTRE , MOOLAKKADAI - 636 303.	Certificate No:	VCPL/CC/23-24/45-003
	Date of Calibration:	05-10-2023
	Date of issue:	06-10-2023
	Date of Receipt:	05-10-2023
	Next cal Date:	04-10-2024

### Details of Device Under Calibration

Description:	SEMI AUTO ANALYZER	Make:	ROBONIK
Range:	-	Model/Type:	PRIESTEST TOUCH
Resolution:	-	Condition of DUC	GOOD
Serial Number	AT0990413RBK	Location:	LABORATORY
ID number	MGPHC/SAA/01	Cal. At:	ONSITE
<b>Environmental Conditions</b>		<b>Calibration Method Used</b>	
Temperature	25.2 °C	National/International standard:	IEC60601-1
Humidity	49 % Rh	Cal procedure no:	-

### Master Details

SL no:	Description	ID No / SL No	Certificate No	Make/Model	Valid Till
1	Electrical Safety Analyzer	6014025	ACCS230581	FLUKE/ESA-612	9-Mar-24

### Electrical Safety Report

S.NO	PARAMETERS	MEASURED VALUES $\mu$ A	ALLOWABLE LIMITS $\mu$ A
1	Earth Leakage	125.6	<5000 $\mu$ A For B,BF,CF
		<b>MEASURED VALUES <math>\mu</math>A</b>	<b>ALLOWABLE LIMITS <math>\mu</math>A</b>
2	Enclosure Leakage	158.9	<5000 $\mu$ A For B,BF,CF

### Remarks:

- 1.DUC is Defined as Device Under Calibration.
- 2.Calibration Frequency decided by the customer.
- 3.The above instrument is working in good condition.

Calibrated By:

  
C.AKASH

(CALIBRATION ENGINEER)



Authorised by:

  
T. SUDHAKAR,  
Technical Manager  
Vanavil Calibrations Pvt Ltd.

## Vanavil Calibrations Pvt. Ltd

(A Unit of Rainbow Bio Medical Engineers)

S2, Rishi Maharaj Apartment, Indra Nagar, 250 Reddiyur Main Road, Narasothipatty, Salem - 636 004.

No. 96,97,98, Anadhamaya Apartments, 2<sup>nd</sup> Floor, KVR Flats, 2<sup>nd</sup> Cross Street, Thangam Nagar, Pallikaranai, Chennai - 600 100.

Cell: 99949 13872 / 99949 23872 Ph: 0427-7965376

✉ vanavilcalibrations@gmail.com 🌐 www.rainbowbiomedicalengineers.com | www.vanavilcalibrations.com