



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JULY 2023



PC-1024

Lab Name

Govt. UPHC, Mayiladuthurai

Lab No

11310

Constituent Group

Chemistry III

Date of Result Entered :

19/07/2023

PT Item

Lyophilized human serum based

Date of Report Published :

07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation)	75	109.71	5.43	5.95	110 mg/dL	0.05	1.37
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation)	66	5.89	4.60	0.27	5.7 mg/dL	-0.70	0.07
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	148	112.78	4.33	4.88	103 mg/dL	-2.00	0.80
SDI Range										
Within -1.00 to +1.00				Excellent.						
Within ± 1.01 to ± 2.00				Good.						
Within ± 2.01 to ± 2.99				Accept with caution. Warning Signal.						
Beyond ± 3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :

Govt. UPHC, Mayiladuthurai
THE MEDICAL OFFICER, MAYILADUTHURAI
Mayiladuthurai
TAMILNADU609001

Paula Christudoss
Medical Officer,
Urban Primary Health Centre,
Mayiladuthurai

Coordinator Contact Details:

Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore