



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - SEPTEMBER 2023



PC-1024

Lab Name Govt. PHC, Viruthanur

Lab No 12378

Constituent Group Chemistry III

Date of Result Entered : 20/09/2023

PT Item Lyophilized human serum based

Date of Report Published : 05/10/2023

| Sl.No | Analyte | Method / Principle Name | Analyzer Name | No of Participants | DV | Participants | | Your Value | SDI | U |
|--------------------------|----------------|---------------------------------------|--|--|--------|--------------|-------|------------|-------|------|
| | | | | | | CV | SD | | | |
| 1 | GLUCOSE III | GOD-POD III | Robonik (Semi automatic / Fully Automatic) | 1870 | 310.86 | 4.82 | 14.98 | 319 mg/dL | 0.54 | 0.69 |
| 2 | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik (Semi automatic / Fully Automatic) | 1852 | 1.29 | 8.14 | 0.10 | 1.28 mg/dL | -0.10 | 0.00 |
| 3 | CHOLESTROL III | CHOD-PAP III | Robonik (Semi automatic / Fully Automatic) | 1783 | 113.62 | 3.80 | 4.32 | 109 mg/dL | -1.07 | 0.20 |
| SDI Range Interpretation | | | | | | | | | | |
| Within -1.00 to +1.00 | | | | Excellent. | | | | | | |
| Within ±1.01 to ±2.00 | | | | Good. | | | | | | |
| Within ±2.01 to ±2.99 | | | | Accept with caution. Warning Signal. | | | | | | |
| Beyond ±3.0 | | | | Unacceptable performance. Action Signal. | | | | | | |

LAB ADDRESS :

Govt. PHC, Viruthanur
CLINICAL LABORATORY
Viruthanur
TAMILNADU625009

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

Pamela Christodoss

Dr. Pamela Christodoss
CMC EQAS Coordinator
Christian Medical College, Vellore

M. S. S. S. S.

கருத்துகள் அல்லாதவை
ஆக ஆரம்ப சுகாதார நிலையம்
விரிவுகூர்ச்சி
மதுரை-9.