

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. UPHC, Mayiladuthurai
 Lab No 11310
 Month June
 Year 2023
 Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)


[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 20/06/2023

Date of Report Published : 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation)	65	287.55	4.50	12.95	298 mg/dL	0.81	3.21
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation)	67	2.58	6.39	0.16	2.5 mg/dL	-0.48	0.04
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	149	114.61	4.05	4.64	110 mg/dL	-0.99	0.76
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						


Medical Officer,
 Urban Primary Health Centre,
 Mayiladuthurai