



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - APRIL 2023



PC-1024

Lab Name: Govt. UPHC, Mayiladuthurai

Lab No: 11310

Constituent Group: Chemistry III

Date of Result Entered: 18/04/2023

PT item: Lyophilized human serum based

Date of Report Published: 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation)	74	118.33	5.77	6.83	101 mg/dL	-2.54	1.59
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation)	65	2.14	7.02	0.15	2.1 mg/dL	-0.27	0.04
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	143	111.22	4.31	4.79	108 mg/dL	-0.67	0.80
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ± 1.01 to ± 2.00				Good.						
Within ± 2.01 to ± 2.99				Accept with caution. Warning Signal.						
Beyond ± 3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :
 Govt. UPHC, Mayiladuthurai
 THE MEDICAL OFFICER, MAYILADUTHURAI
 Mayiladuthurai
 TAMILNADU609001

Medical Officer,
 Urban Primary Health Centre,
 Mayiladuthurai

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore